

Nederlands Meetinstituut

Member State
The Netherlands

OIML Certificate N° R60/2000-NL-02.16 Revision 1
Project number 206059
Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1, Dordrecht
Person responsible: P.P.M. van Enkevort

Applicant

Name: Hottinger Baldwin Messtechnik GmbH
Address: Im Tiefensee 45
D-64293 Darmstadt
Germany

Manufacturer of the certified pattern

Name: Hottinger Baldwin Messtechnik GmbH
Address: Im Tiefensee 45
D-64293 Darmstadt
Germany

Identification of the certified pattern

Type : PW15../..
Fraction : $P_i = 0.7$
Temperature limit : -10 °C / 40 °C

Maximum capacity (E_{max})	7.5, 15, 20, 30, 50, 75, 100, 150 and 200 kg					
Accuracy Class	D	C				
Maximum number of load cell intervals (n)	1000	1000	1500	2000	2500	3000
Ratio of minimum LC Verification interval $Y = E_{max} / V_{min}$	2500	7500			7500	10000 15000


Nederlands Meetinstituut
Hugo de Grootplein 1
3314 EG Dordrecht
Telephone +31 78 6332332
Telefax +31 78 6332309

NMI B.V.
(Chamber of Commerce no.27.228.701)

Subsidiary companies:
NMI Van Swinden Laboratorium B.V. (27.228.703)
NMI Certin B.V. (27.233.418)
Verispect B.V. (27.228.700)

This document is issued under the provision that NMI, B.V. nor its subsidiary companies accept any liability.

Reproduction of the complete document is allowed. Parts of the document may only be reproduced after written permission



Nederlands Meetinstituut

Member State
The Netherlands

OIML Certificate N° R60/2000-NL-02.16 Revision 1
Project number 206059
Page 2 of 2

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report, the test certificate and the description with number TC5384 and the appertaining documentation folder), with the requirements of the following Recommendation(s) of the International Organisation of Legal Metrology (OIML):

R60
edition 2000 (E)
for accuracy class C and D

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test report:
N° R60/1991-NL-98.19A, that includes 40 pages
N° R60/1991-NL-98.19B, that includes 37 pages
N° R60/2000-NL-02.16, that includes 37 pages.

This revision replaces the earlier version.

The issuing authority
P.P.M. van Enckevort
Manager Certification Delft
1A

10 June, 2002

The OIML member
G.J. Faber

10 June, 2002

*
**

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.