

Nederlands Meetinstituut

Member State
The Netherlands

OIML Certificate N° R60/1991-NL-98.07
Project number 10086508
Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1, Dordrecht
Person responsible: A.J. Nederlof

Applicant

Name: HBM Wägetechnik GmbH
Address: Im Tiefen See 45
D-64293 Darmstadt
Germany

Manufacturer of the certified pattern

Name: HBM Wägetechnik GmbH
Address: Im Tiefen See 45
D-64293 Darmstadt
Germany

Identification of the certified pattern

Type : PW12
Fraction : $P_i = 0.7$

Maximum capacity (E_{max})	150, 200, 250, 300, 500 and 600 kg					
Accuracy Class	D	C				
Maximum number of load cell intervals (n)	1000	1000	1500	2000	2500	3000
Ratio of minimum LC Verification interval $Y = E_{max} / V_{min}$	30000					
Maximum platform size according OIML R76	800 X 800 mm					

Nederlands Meetinstituut
Hugo de Grootplein 1
3314 EG Dordrecht
Telephone +31 78 6332332
Telefax +31 78 6332309

NMI B.V. (Chamber of Commerce Haaglanden
No.27228701)


Subsidiary companies:
NMI Certin B.V. (27233418)
NMI Van Swinden Laboratorium B.V. (27228703)
NMI International B.V. (27239176)

This document is issued under the provision that NMI. B.V. nor its subsidiary companies accept any liability.

Reproduction of the complete document is allowed. Parts of the document may only be reproduced after written permission



QUALIFIED
BY STERLAB
Reg. nr. L 029



Nederlands Meetinstituut

Member State
The Netherlands

OIML Certificate N° R60/1991-NL-98.07
Project number 10086508
Page 2 of 2

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report, the test certificate and the description with number TC5259 and the appertaining documentation folder, with the requirements of the following Recommendation(s) of the International Organisation of Legal Metrology (OIML):

R60
edition 1991 (E)
for accuracy class C and D

This certificate relates only to the metrological and technical characteristics of the pattern of the Load cell concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test report:
N° R60/1991-NL-98.07, that includes 39 pages.

The issuing authority
A.J. Nederlof
110

08 April 1998

The OIML member
G.J. Faber

08 April 1998

*
**

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.