

Nederlands Meetinstituut

Member State
The Netherlands

OIML Certificate N° R60/2000-NL-00.09

Project number 10106785

Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1, Dordrecht
Person responsible: Ms. B. van Broekhoven

Applicant

Name: Hottinger Baldwin Messtechnik GmbH
Address: Im Tiefen See 45
D-64293 Darmstadt
Germany

Manufacturer of the certified pattern

Name: Hottinger Baldwin Messtechnik GmbH
Address: Im Tiefen See 45
D-64293 Darmstadt
Germany

Identification of the certified pattern

Type : PW6K. .../.. K
Fraction : $P_i = 0.7$

Maximum capacity (E_{max})	3, 5, 10, 15, 20, 30 and 40 kg
Accuracy Class	C
Maximum number of load cell intervals (n)	3000
Ratio of minimum LC Verification interval $Y = E_{max} / V_{min}$	20000
Maximum platform size, according OIML R76	300 x 300 mm
Humidity classification SH	3 and 5 kg
Humidity classification CH	10, 15, 20, 30 and 40 kg

Nederlands Meetinstituut
Hugo de Grootplein 1
3314 EG Dordrecht
Telephone +31 78 6332332
Telefax +31 78 6332309

NMI B.V. (Chamber of Commerce Haaglanden
No.27228701)


Subsidiary companies:
NMI Certin B.V. (27233418)
NMI Van Swinden Laboratorium B.V. (27228703)
NMI International B.V. (27239176)

This document is issued under the provision that
NMI, B.V. nor its subsidiary companies accept
any liability.

Reproduction of the complete document is
allowed. Parts of the document may only be
reproduced after written permission



QUALIFIED
BY STERLAB
Reg. nr. L 029



Nederlands Meetinstituut

Member State
The Netherlands

OIML Certificate N° R60/2000-NL-00.09
Project number 10106785
Page 2 of 2

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report, the test certificate and the description with number TC5159 and the appertaining documentation folder), with the requirements of the following Recommendation(s) of the International Organisation of Legal Metrology (OIML):

R60
edition 2000 (E)
for accuracy class C and D


This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test reports:


N° R60/1991-NL-97.30, that includes 40 pages;
N° R60/2000-NL-00.09A, that includes 40 pages;
N° R60/2000-NL-00.09B, that includes 12 pages.

The issuing authority
Ms. B. van Broekhoven
Manager Certification Delft



22 June, 2000

The OIML member
G.J. Faber



22 June, 2000

*
* *

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.