

Nederlands Meetinstituut

Member State
The Netherlands

OIML Certificate N° R60/1991-NL-98.13

Project number 10089037

Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1, Dordrecht
Person responsible: A.J. Nederlof

Applicant

Name: HBM Inc.
Address: 19 Bartlett Street
Marlboro, MA 01752
USA

Manufacturer of the certified pattern

Name: HBM Inc.
Address: 19 Bartlett Street
Marlboro, MA 01752
USA

Identification of the certified pattern

Type : SP4
Fraction : $P_1 = 0.7$

Maximum capacity (E_{max})	7, 10, 15, 20, 30, 50, 75 and 100 kg
Accuracy Class	C
Maximum number of load cell intervals (n)	3000
Ratio of minimum LC Verification interval $Y = E_{max} / V_{min}$	14000

Nederlands Meetinstituut
Hugo de Grootplein 1
3314 EG Dordrecht
Telephone +31 78 6332332
Telefax +31 78 6332309

NMI B.V. (Chamber of Commerce Haaglanden
No.27228701)


Subsidiary companies:
NMI Certin B.V. (27233418)
NMI Van Swinden Laboratorium B.V. (27228703)
NMI International B.V. (27239176)

This document is issued under the provision that
NMI, B.V. nor its subsidiary companies accept
any liability.

Reproduction of the complete document is
allowed. Parts of the document may only be
reproduced after written permission



QUALIFIED
BY STERLAB
Reg. nr. L 029



Nederlands Meetinstituut

O Member State
The Netherlands

OIML Certificate N° R60/1991-NL-98.13
Project number 10089037
Page 2 of 2

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report, the test certificate and the description with number TC2485 and the appertaining documentation folder, with the requirements of the following Recommendation(s) of the International Organisation of Legal Metrology (OIML):

R60
edition 1991 (E)
for accuracy class C

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test reports:

N° R60/1991-NL-95.11, that includes 38 pages;
N° R60/1991-NL-95.15, that includes 35 pages;
N° R60/1991-NL-98.13, that includes 12 pages.

The issuing authority
A.J. Nederlof

110

07 May 1998

The OIML member
G.J. Faber

07 May 1998

*
* *

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.